


USER GUIDE



Model SL355

Noise Dosimeter, Datalogger, and Sound Level Meter With PC Interface



Copyright © 2012 Extech Instruments Corporation (a FLIR company)
All rights reserved including the right of reproduction in whole or in part in any form.

1. Table of Contents

1.	Table of Contents	2
2.	Introduction	3
3.	Features	4
4.	Meter Description	4
5.	Display Description	5
6.	Preparation	6
6.1	Getting Started	6
6.2	Battery Installation, Replacement, and Disposal	6
6.3	Batter Status Symbols	6
6.4	Microphone Installation	6
7.	Calibration Procedure	7
8.	Dose Testing and Datalogging	8
8.1	Overview of Dose Testing and Datalogging	8
8.2	Running Dose Tests and Datalogger	8
8.3	Reviewing Dose Test Setups and Measurement results	9
8.4	Review Stored Datalogger Data	9
8.5	Data Mode - Menu Table	10
8.6	Data Mode - Meter Screen Shots	11
8.6.1	Select Dose Meter Setup from Memory	11
8.6.2	Measure and Store Dose Data	11
8.6.3	Screen Shots for Recalling and Reviewing Dose Data	14
9.	Sound Level Meter (SLM) Mode	15
9.1	SLM Menu Table	15
9.2	SLM Screen Shots	16
10.	Meter Setup - Software Method	18
11.	Meter Setup - Manually	18
11.1	SETUP Mode	18
11.2	SETUP Mode - Menu Table	18
11.3	SETUP Mode - Screen Shots	19
11.4	OPTION Mode	21
11.5	OPTION Mode - Menu Table	21
11.6	OPTION Mode - Basic Screen Shots	23
11.7	OPTION Mode - Set the Time and Date	24
11.8	OPTION Mode - Editing	25
12.	Clear the Meter's Memory	27
13.	Specifications	28
14.	Maintenance	31
15.	Software Installation and Operation	31
16.	Glossary of Terms	32

2. Introduction

Congratulations on the purchase of the Model SL355. The SL355 is a combination Noise Dosimeter, Datalogger, and Sound Level Meter (SLM). The SL355 can be configured manually via push-button navigation of the menu system or through the PC interface using the supplied software (recommended for ease of use). Refer to the dedicated software User Guide on the supplied CD-ROM for detailed instructions on using the software program for downloading data, configuring the SL355 meter, and remotely controlling the meter.

The **SL355 Noise Dosimeter** is designed to test noise exposure and provide noise surveys for compliance with OSHA, MSHA, DOD, ACGIH, and ISO standards. The SL355 monitors accumulated noise exposure and measures frequency weighted noise exposure and peak sound level simultaneously for % dose and TWA (time weighted average or sound exposure in Pa²h).

Up to twenty (20) noise surveys can be recorded and up to five (5) automated dose tests can be programmed up to 30 days in advance. Each test includes a 'repeat' feature where tests can be rerun a pre-programmed number of times. Criterion Level, Threshold, Exchange rate and other parameters are user-programmable. Noise Survey results can be viewed on the meter's display and downloaded using the supplied PC software.

The **SL355 Sound Level Meter (SLM)** displays real-time sound pressure level (SPL) readings along with LEQ, SEL, and elapsed time as a typical type 2 SLM would. The SLM has selectable measurement ranges: 60 to 130 & 70 to 140dB, 'A' and 'C' frequency weightings, Linear 'Z', and 'C' Peak weightings, Fast & Slow response time settings, and an adjustable energy exchange rate (Q).

The **SL355 Datalogger** has a built-in real-time clock for date/time stamped data records. The datalogger can store up to 14,400 readings in the meter's memory. Sampling time (rate) is programmable from 1 second per reading up to 1 hour per reading. Logged readings are downloaded using the supplied PC software.

This device has been calibrated and tested before shipment. With proper care, years of reliable service can be expected.

Conventions of this User Guide

The SL355 is a sophisticated device with a variety of features. This User Guide is intended to optimize the user experience with regard to ease of use and convenience.

To accomplish this, most sections of this User Guide include instructional text, a Menu table, and a section that illustrates the actual "Screen shots" seen by the user while programming the meter.


Be sure to review all three of these approaches (instructional text, menu tables, and screen shots) and determine if one, or a combination of the three, work best for the application at hand. If questions arise please contact our technical support staff.

3. Features

- Combination Dose Meter, Datalogger, and Sound Level Meter.
- User-defined measurement configurations. Selectable Criterion Level, Exchange Rate, Threshold, Frequency Weighting, and Response Time settings.
- Completely configurable using the supplied PC software (recommended).
- Twenty (20) internal memory locations with a total capacity of 14,400 readings.
- Download Dose, TWA, and Datalogger data to a PC using supplied PC software.
- Real-time clock and date/time stamp datalogging capability.
- Keypad lock protects against inadvertent meter operation.
- Included USB PC interface and Windows™ configuration and datalogging software.
- Kit includes a microphone, microphone clip, microphone windscreen, belt clip, batteries, protective case, keypad cover, user guide, software CD-ROM, and USB cable.

4. Meter Description

1. Microphone input jack
2. Calibration potentiometer for 114db adjustment
3. LCD Display
4. ON-OFF and MENU button
5. Left arrow button
6. Down arrow button
7. USB PC interface jack
8. Right arrow button
9. Up arrow button
10. RUN/PAUSE and RETURN button















Keyboard cover:



Notes: Battery compartment and belt clip are located on rear of meter. Microphone lapel clip, calibration screwdriver, microphone, windscreen, and other kit items are not pictured.

5. Display Description


1.  Over range indicator
2.  Under range indicator
3.  Flashing: Timer standby; Solid: Recording
4.  Recording Pause indicator
5.  'Smart' navigation buttons
6. **SPL** (Sound Pressure Level)
PK (Peak mode: 'C' frequency weighting)
PKZ (Peak mode: Linear 'Z' weighting)
DOSE (%dose)
LEQ (Equivalent Continuous Sound Level)
TWA (Time Weighted Average)
SEL (Single Event Noise Exposure Level)
7. **S** Slow (response time weighting)
F Fast (response time weighting)
C 'C' frequency weighting
A 'A' frequency weighting
8. **dB%** Percent or decibel readout units
9.  Battery status
10. Primary reading, date/time display, or menu items for the programming modes
11. Mode Icons:
 Sound level meter mode
 Dosimeter/Datalogger (when flashing, Data is recording and the keypad is locked; when solid, the meter is waiting for the Timer to countdown before recording can begin)
 Setup mode of operation
 Option mode of operation
12.  Headphone icon indicates that a SPL of over 115db has occurred
13. **BEG** Start test time
END Stop test time
DUR Test duration



6. Preparation


6.1 Getting Started

1. Power the instrument by momentarily pressing the power/menu button at the upper left side of the keypad.
2. When the unit is turned on, the model number (355) appears briefly on the left side of the LCD and the firmware version (1.1, for example) appears on the right side.
3. The meter defaults to the SLM mode (SLM icon is shown) where the main display indicates the sound pressure level (SPL) in db, the A/C frequency weighting setting, the F/S response time setting, and the battery status icon (an up or down arrow will be visible if the current reading is above or below the selected range respectively). See screen shot diagram.
4. If the meter will not switch on, check that the batteries are installed and are fresh.
5. To turn the meter OFF, press and hold the power/menu button for three (3) seconds.
6. The SL355 can be used as a Sound Level Meter or a Dosimeter/Datalogger.
7. Note that the four arrow buttons are 'smart' navigation keys. For any given meter mode only the active arrow buttons appear on the LCD.



6.2 Battery Installation, Replacement, and Disposal


1. Slide the rear battery cover off of the meter.
2. Remove the old batteries (if applicable) and insert three (3) 'AAA' 1.5V batteries. Ensure that the battery orientation and polarity are correct. Do not mix fresh and partially used batteries.
3. Carefully re-assemble the meter before use.



The end user is legally bound (EU Battery ordinance) to return all used batteries, disposal in the household garbage is prohibited! Bring used batteries / accumulators to collection points in your community or wherever batteries / accumulators are sold! Disposal: Follow the valid legal stipulations in respect of the disposal of the device at the end of its lifecycle

6.3 Battery Status

The battery status icon appears as empty when the battery is depleted. The battery symbol indicates the remaining capacity in numbers of bars or segments, as follows:




6.4 Microphone installation

Insert the microphone plug into the connector on the top of the meter as shown in accompanying diagram. Turn clockwise to secure in place; be careful not to over-tighten.




7. Calibration Procedure

1. A standard 114db acoustic calibrator is required (as shown in diagram below).
2. Insert the meter's microphone carefully into the insertion hole of the calibrator (as shown).



3. Switch the calibrator ON and set its output amplitude to 114 db.
4. Switch the meter ON.
5. Access the Calibration mode (CL) by repeatedly pressing the MENU button until the **CL XX.X** display is shown and the **SLM**, **DATA**, **SETUP**, and **OPTION** mode icons are all switched off (refer to accompanying screen shot illustration).



6. Adjust the meter's potentiometer (recessed at top of meter) using the supplied adjustment tool until the display reads 114db, matching the 114db output level of the calibrator.
7. Press the RUN/PAUSE/RETURN button to save the calibration.

8. Dose Testing and Datalogging

8.1 Overview

The dosimeter noise survey and datalogging functions operate at the same time. For example, while the %dose and TWA (time weighted average) values are being calculated and displayed (for the dose meter's noise survey) individual SPL readings are being stored by the datalogger.

Noise surveys and logged data are stored in 'RUNS' which are memory locations (1 to 20). Note that the meter can collect up to 14,400 readings total, spread out over the 20 RUN memory locations. The 'RUN' parameter is located in the DATA menu. The DATA Mode Menu table and a Screen-Shot section are provided directly after the instructional text below:

8.2 Running the Dose Meter Noise Survey and Datalogger

1. Use the supplied software program (recommended) to configure the meter's noise survey and datalogging parameters (a software User Guide is provided on the supplied CD-ROM), or manually configure the meter following the instructions provided in the SETUP and OPTION mode sections of this guide. For compliance, the noise survey parameters must match the requirements of the regulation/standard in question.
2. In the SETUP mode (or using the supplied software) set the Criterion Level, Exchange Rate, Threshold, Frequency Weighting, Response time, Range, and Peak parameters. Note that if a configuration has already been stored, it can be recalled in the SETUP mode. Reference Section 11.3 for SETUP Mode details.
3. In the OPTION mode (or using supplied software) set the Date/Time, Test Duration, Test Begin (BEG) time, Test End time, Sample rate (logging period), Repeat, Pair, and Timer functions. Reference Section 11.5 for OPTION Mode details.
4. For the entirety of the test, the microphone should be clipped to the test person's lapel and the meter should be clipped to the test person's belt.
5. To run the dosimeter/datalogger the meter must be in the DATA Mode of operation. Press the Menu button to select the DATA Mode (if not already selected).

TIMER Notes:

- The Timer setting is critical for determining whether the test begins immediately (Timer OFF) or at a programmed time (Timer ON).
- With the TIMER set to OFF, recording can begin immediately by pressing the RUN/PAUSE button. When the meter is recording, the REC icon will appear solid and the DATA icon will flash.
- With the TIMER set to ON, and with the Beginning and End times programmed, after RUN/PAUSE is pressed recording will NOT start until the programmed Begin (BEG) time is reached. The REC icon will be flashing (indicating that the meter is in "standby" waiting for the Begin time); the DATA icon will be on solid.
- Up to 5 Begin/End times can be pre-set and tests can be programmed up to 30 days in advance.

DUR (DURATION) Notes:

- DUR allows the user to select a specific RUN Duration time.
 - When a duration period is set by the user to anything other than OFF, it takes precedence over the Timer setting so that any preset Begin and End times will be ignored. DUR is available in the OPTION mode menu.
6. While testing, the meter's DATA menu can be navigated (see DATA Menu table and Screen Shots below) as long as the keypad is unlocked. This allows the user to view the current setup, % dose, and TWA while the test is running.
 7. The arrow keys are locked automatically when recording begins. Press ◀ and ▼ simultaneously to unlock the keypad. The DATA icon will cease flashing, indicating that the keypad is unlocked and ready for navigation. To re-lock the keypad, press ▲ and ▶ simultaneously.
 8. To pause a recording, press RUN/PAUSE. The pause icon || will appear. Press again to resume.
 9. To stop the recording if the END time has not been automatically set, first unlock the keypad by pressing ◀ and ▼ simultaneously and then press the MENU key. The display will show STOP and will count down from 3 to zero; press the RUN/PAUSE button before zero is reached to successfully stop the recording.
 10. Switch the meter off to conserve battery power. To switch the meter off, press and hold the menu button until the unit switches off.
 11. %Dose and TWA calculations are viewable using the PC interface or by navigating the DATA mode (refer to the DATA Menu table and Screen Shots below). Individually logged readings are retrievable via the PC interface only. Follow the directions in the Software User Guide supplied on the enclosed CD-ROM.

Testing notes:

- Do not shout into the microphone or strike the microphone as this will affect the test.
- To obtain reliable data, the user's activities must reflect a typical workday.

8.3 Reviewing Noise Survey Setups and Measurement Results

1. Note that the PC interface can be used to review setups and measurement results. Refer to the Software User Guide on the supplied CD-ROM for instructions. To work manually, follow the steps below.
2. Switch the meter ON; the meter will be in the SLM mode. Press MENU to access the DATA mode.
3. Use ▶ to scroll to the 'RUN' display. Now use ▼▲ to select the desired RUN (memory location).
4. When the desired RUN is selected, use ◀▶ to scroll the setup data and noise survey results as shown in the DATA Menu table and Screen Shots below.

8.4 Reviewing Stored Datalogger Data


The stored datalogger readings can only be viewed by downloading the data to a PC using the supplied cable and PC software program. Refer to the Software User Guide on the supplied CD-ROM for detailed instructions.

8.5 DATA Mode Menu Table




Display	Description	Display	Description
HH:MM:SS (Use ▼ to scroll items below)	Current Time	HH:MM:SS (Use ► to scroll items below)	Current Time
DD:MM:YY	Current Date	BEG	Datalogger Begin Time
SPL	Sound Pressure Level	rUN [†]	Current memory location (01-20)
		DUR	Test Duration
		USr (selected setup)	View %DOSE and TWA using ▲ and ▼ buttons
		60 to 130db 70 to 140db	Measurement Ranges
		Er	Exchange rate (Q) in db
		tHr	Threshold in db
		Crit	Criterion level in db
		LP	Logging Period (mm:ss) from 1 sec. to 1 hour
[†] Scroll the RUN parameter list using the ▼ ▲ buttons to locate the EPty XX parameter which indicates the number of empty memory locations or runs remaining (00 to 20); if recording is attempted with all memory locations full, the meter will display 'EPty00' indicating that there are no (zero) empty storage locations remaining (data must then be cleared from the meter before continuing; use the OPTION menu to clear data).			

8.6 SCREEN SHOTS for Dose Meter Operation







8.6.1 Select a Dose Meter Setup from Memory

Operation	Press	Display Shows	Comments
Change the Mode of Operation to SETUP	MENU repeatedly		Select SET UP to choose a stored setup configuration. The first display always shows the currently loaded setup, (for example USr, as in the screen shot at left). If no user setups are stored, the up/down arrow buttons will not appear on the LCD.

8.6.2 Measure and Store Dose Data












Operation	Press	Display Shows	Comments
1. Change Operating Mode	MENU repeatedly		The first screen shows the time. Time format is hh:mm:ss. Date format is dd-mm-yy (▼) SPL display (▼)
2. Start Run	RUN / PAUSE		When the run TIMER is ON with Beginning and End times set, the meter waits, showing the REC icon flashing, and the time remaining before the run is set to begin. When the run TIMER is ON, the meter waits until the next whole clock minute, then starts the run. When the run TIMER is OFF, the meter waits until the next whole clock second, then starts the run.
3. Test is Running			REC icon indicates that the meter is now storing dose data in the first available data storage location.

Operation	Press	Display Shows	Comments
4. Inspect the setup for the current RUN	▶ To Scroll	<p>RUN Begin Time (dd-hh:mm)</p> <p>Logging Period (mm:ss)</p> <p>Criterion Level 90db</p> <p>Threshold 70db</p> <p>Exchange Rate 3db</p> <p>Storage Location for RUN (04)</p> <p>RUN Duration (hh:mm:ss)</p> <p>Selected user setup (for example, ISO with 90db Criterion Level)</p> <p>Measurement Range</p>	

5. Inspect data from the current RUN	▼ several times	 <p>User Setup (for example, ISO with 90db Criterion Level) Percent Dose (Dose %) Sound Exposure (Pa²h)</p>	Some displays offer the use of the ▲ and ▼ keys to show other parameters. The display updates while the RUN is in progress.
6. Pause	RUN / PAUSE		The pause icon () is shown. Press this key again to resume the RUN.
7. Lock Keypad	▶ and ▲ at the same time		To lock the keys to prevent unauthorized tampering, The DATA icon will be flashing. The keys and display cannot be locked while the meter is paused.
8. Unlock Keypad	◀ and ▼ at the same time		To unlock the keys and display to restore normal operation.
9. Stop the RUN	MENU		A 3 second countdown will start. This command must be confirmed immediately (before the counter reaches zero) by pressing the RETURN button.
10. Confirm Stop	←		The RUN is now complete.

8.6.3 Screen Shots for Recalling and Reviewing Dose Data

Note that stored data is identified by date (day) and measurement start time (hh:mm) and that up to 20 RUNS can be stored simultaneously.

Operation	Press	Display Shows	Comments
1. Switch the meter ON			The meter always powers up in the Sound Level Meter mode.
2. Change the Operating Mode	MENU		Selecting DATA mode.
3. Enable Data Recall			Display the Start Date & Time for the most recently stored RUN (dd-hh:mm).
4. Review Stored RUNS	 repeatedly		Select RUN by Start Date & Time.
5. Review Stored RUN			Shows where data from the selected run is stored
6. Review Stored Data			The stored data can be reviewed as described in steps 4 and 5 of section 8.6.2 above.

9. Sound Level Meter (SLM) Mode

Use the MENU button to access the SLM mode; the **SLM** icon will appear on the meter's LCD when the SLM mode is accessed. The Elapsed Timer will start automatically and the meter will begin displaying the sound pressure level (SPL) measurements in db (decibels); use the RUN/PAUSE button to pause and resume the timer.

In SLM mode the meter operates as a standard 'type 2' sound level meter, measuring and displaying sound pressure levels from 60 to 130db or 70 to 140dB (the range is selectable as described later in this procedure).

Read the measured Sound Pressure Level (SPL) on the LCD. For Sound Level measurements that fall below the specified range, a down arrow icon will appear on the display; for readings above the specified range, an up arrow icon will appear on the LCD. For readings > 115dBA, the headphone icon appears. For readings > 140dBA, the Peak Detector (PK icon) appears along with the 'C' or 'Z' icon depending on the selected Peak weighting mode ('C' for 'C' frequency weighting and 'Z' for flat or linear response where no frequency weighting is applied).

In SLM mode the user can also view SEL (Single Event noise exposure Level) and L_{eq} (Equivalent continuous sound pressure Level) readings (refer to the SLM menu for navigation).



Refer to the Menu table and Screen Shots provided below illustrative instructions. In the Menu table below, the user can edit the parameters shown on the right side of the thick black vertical line. When a parameter is displayed, press and hold RUN/PAUSE until the parameter's setting flashes. Press ▲ or ▼ to change the setting and then press RUN/PAUSE again to save the change.

9.1 SOUND LEVEL METER (SLM) Menu Table

Display	Description	Display	Description
SPL (Use ▼ to scroll items below)	Sound Pressure Level in db	SPL (Use ► to scroll items below)	Sound Pressure Level in db
DUR	Test Duration	60 to 130db or 70 to 140db	Measurement Range
LEQ	Equivalent Continuous Noise Level	F or S	Time Weighting Fast/Slow
SEL	Sound Exposure Level	Er (3, 4, 5 or 6 db)	Exchange rate (Q)
PK	Peak SPL for C or Z weighting	A or C	Frequency weighting
Clr N	Clear (Erase data); Use PAUSE/RUN button to erase RUN data ('Y' for Yes or 'N' for No)	PK-C or PK-Z	'Z' (linear) or 'C' frequency weighting for Peak SPL

9.2 Screen Shots for SOUND LEVEL METER Mode

As mentioned, in SLM mode the meter operates as a Type 2 sound level meter. Note that data cannot be logged while the meter is in the SLM mode.

Operation	Press	Display Shows	Comments
1. Switch meter ON	ⓘ		Start from the Sound Level Meter mode.
2. Review SLM Measurement	▼ To Scroll	 <p>Sound Pressure Level A-Frequency Weighting Fast Time Weighting</p> <p>Press RETURN (RUN/PAUSE) button to</p> <p>Max. Peak (L_{LPK} or L_{CPK}) 'Z' for Linear 'C' for C-Weighting</p>	

Operation	Press	Display Shows	Comments
3. Inspect and edit the settings	<p>▶ To scroll</p> <p>← Press to edit</p> <p>▼▲ To change setting</p> <p>← Press to confirm</p>	<p>Press Return (RUN/PAUSE) to edit the Range</p> <p>Press Return (RUN/PAUSE) to edit Max. Peak</p> <p>Press Return (RUN/PAUSE) to edit Fast/Slow Response Time</p> <p>Press Return (RUN/PAUSE) to edit A/C Frequency weighting</p> <p>Press Return (RUN/PAUSE) to</p>	
4. Editing example	←	<p>Current range is flashing</p> <p>Store the range setting</p> <p>Selected range is flashing</p>	
5. Switch OFF	<p>ⓘ</p> <p>press 3 seconds</p>	Meter display switches OFF	

10. Meter setup using supplied software (Preferred Method)


To configure the meter using the supplied software:

1. Install the supplied software onto a PC.
2. Connect the meter to the PC using the supplied USB cable.
3. Run the software program.
4. Refer to the Software User Guide provided on the supplied CD-ROM for detailed instructions on the use of the software program.

11. Configuring the Meter Manually

11.1 SETUP MODE

In SETUP mode, the parameters listed in the SETUP Mode Menu table and screen shot section below can be configured. The first display in the SETUP mode is always the currently selected setup configuration as shown in the accompanying diagram.






1. Switch the meter ON using the power button at the upper left side of the keypad.
2. Note that the four arrow buttons are 'smart' navigation keys. For any given meter mode only the *active* arrow buttons appear on the LCD.
3. Switch to the Set Up mode using the MENU button until **SETUP** is displayed.
4. Press ◀ or ▶ to step through the parameters shown in the menu table below.
5. To change a setting of a parameter in the menu below:
 - Press and Hold the RUN/PAUSE button when the desired parameter is shown (the setting value will begin flashing indicating that it is ready to be edited).
 - Press ▲ or ▼ to select the desired value.
 - When the desired value is selected, momentarily press the **RUN/PAUSE** button again to save the setting (the value will stop flashing):

11.2 SETUP Mode - Menu Table

Icon	Description
Use ◀ or ▶ to scroll items below. Press and hold the RUN/PAUSE to edit a parameter.	
60 to 130db or 70 to 140db	Measurement Ranges
F or S	Fast or Slow Response Time
Er	Exchange rate (3, 4, 5, or 6 db) also known as 'Q'
A or C	Frequency weighting
PK-C or PK-Z	Peak SPL for C -weighting or Z -weighting (linear)
tHr	Threshold level in db (70 to 90dB in 1 dB steps)
Crit	Criterion level in db (80, 84, 85 or 90dB)

11.3 SETUP Mode – Screen Shots

Operation	Press	Display Shows	Comments
1. Switch the meter ON	ⓘ		The meter powers up in the Sound Level Meter mode.
2. Change Mode	MENU several times		Selecting SET UP mode. The first display shows the name of the currently loaded Setup.
3. Inspect and edit settings	<p>▶ To scroll</p> <p>← Press to edit</p> <p>▼▲ To change setting</p> <p>← Press to confirm</p>	 <p>Press Return (RUN/PAUSE) to edit the Range</p> <p>Press Return (RUN/PAUSE) to edit Fast/Slow Response Time</p> <p>Current User Setup Main screen</p> <p>Press Return (RUN/PAUSE)</p> <p>Press Return (RUN/PAUSE)</p> <p>Press Return (RUN/PAUSE)</p> <p>Press Return (RUN/PAUSE)</p>	

4. Editing example



Changing Criterion Level, for example



The current Criterion Level will be flashing

OR



The stored Criterion Level




The new, selected Criterion Level will be flashing

11.4 OPTION MODE

The OPTION menu below lists the parameters that can be viewed and/or configured. Instructions for navigation and use are provided in the menu table and screen shots below.

1. Switch the meter ON using the power/menu button.
2. Use the MENU button to access the OPTION mode. **OPTION** will be shown when the meter is in the OPTION mode and the current time of day will be displayed (hh:mm:ss).
3. Press ◀ or ▶ to scroll the parameters in the OPTION mode. Refer to the Option Menu table and Screen Shots below for further details.






11.5 OPTION Mode - Menu Table





Display (Use ▶ to scroll items below)	Description
HH:MM:SS	Current Time
DD:MM:YY	Current Date
bAt	Battery voltage level (new alkaline batteries ~4.6V)
tlmr	TIMER. Press RUN/PAUSE to toggle the Timer ON/OFF. Press ▶ to skip to the CLrxx parameter (Clear memory) discussed at the bottom of this menu table. With the TIMER set to ON only, use ▼ to scroll through the test setup items below:
LP	Logging Period (mm:ss). Select a datalogging sampling rate from one second up to 60 minutes. Press and hold RUN/PAUSE for 3 seconds and the time display will flash. Press ▲ or ▼ to change the period and then press RUN/PAUSE again. Press ▼ to move to the next parameter.
DUR	Test duration programmable in hh:mm format. Press and hold RUN/PAUSE for 3 seconds; the time display will flash. Use ▲ or ▼ to change the time (or to select OFF). When the desired test duration is displayed, press RUN/PAUSE again. If the DUR is set to OFF, the following parameters will be available; press ▼ to access them.

rEPt	<p>The REPT (Repeat) function allows the test to be repeated a pre-programmed number of times (OFF, 0 to 19).</p> <p>Press and hold RUN/PAUSE for 3 seconds; the display will flash.</p> <p>Use ▲ or ▼ to change the setting and then press RUN/PAUSE.</p> <p>Press ▼ to move to the next parameter.</p>
PAIr	<p>Tests 'pairs' are pairs of Start (BEG) and End times. When a test pair is active, a small 'L' appears next to the 'P' in 'Pair'.</p> <p>To disable the test pair, press RUN/PAUSE and the 'L' will switch OFF.</p> <p>Press ▼ to move to the next parameter.</p>
BEG	<p>Test Start (BEGIN) time (HH:MM:SS);</p> <p>If the TIMER is ON, tests will begin at the BEGIN time after the RUN/PAUSE button is pressed.</p> <p>To change the Start time, press RUN/PAUSE, a pair of digits will flash; Press ▲ or ▼ to change the flashing digits.</p> <p>Press ◀ or ▶ to select another pair of digits. When done, press RUN/PAUSE to store the new Start time.</p> <p>Press ▼ to move to the next parameter.</p>
END	<p>Test END time (HH:MM:SS)</p> <p>To change the Test END time, press RUN/PAUSE, a pair of digits will flash; Press ▲ or ▼ to change the flashing digits.</p> <p>Press ◀ or ▶ to select another pair of digits. When done, press RUN/PAUSE to store the new End time.</p> <p>Use the MENU button to change mode of operation.</p>
CLrxx	<p>Access this parameter from the TIMER parameter earlier in this menu.</p> <p>Clear memory location (Run: 01-20); Press ▼ to access the parameter below or press ▶ to return to the current time display</p>
SUreXX	<p>Are you sure? Press RUN/PAUSE to erase the current memory location.</p> <p>Use the MENU button to change the mode of operation.</p>

11.6 OPTION Mode - Basic Screen Shots






Operation	Press	Display Shows	Comments
1. Switch the meter ON	ⓘ		Start from the Sound Level Meter mode.
2. Change Mode	MENU several times		Current Time.
3. Scroll options	▶ To scroll		<p>The display cycles through the remaining options:</p> <ul style="list-style-type: none"> Current Time–hh:mm:ss Current Date–dd:mm:yy Battery status: voltage Timer Status Data Clear – Use this option with caution – ALL DATA CAN BE DELETED!




11.7 Set the Time & Date in OPTION Mode

Operation	Press	Display Shows	Comments
1. Time Option			Continued from Step 2 in Section 11.6
2. Enable Time Change	←		Cursors are enabled and the HOURS entry blinks to indicate that it can be edited.
3. Change the entry	▲ or ▼		Use ► and ◀ to highlight an entry. Use ▲ and ▼ to change the entry.
4. Save the Changes	←		The new time setting is stored.

Repeat this process to set the date. Note that the date format is dd-mm-yy.



11.8 OPTION Mode - Editing

Operation	Press	Display Shows	Comments
1. Select Timer Option			Continuing from Section 11.6
2. Switch Timer ON	←		The TIMER is now ON, therefore the programmed Start and Stop times will be used (unless Duration DUR settings are enabled as shown in step 4 below).
3. Logging Period setting	▼		LP allows the user to select a Logging Period (data sample rate) Press ← to enable, use ▲ and ▼ to select from the following Logging periods: 1, 2, 5, 10, 15, 20, 30 seconds, 1, 2, 5, 10, 15, 20, 30, 60 minutes.
4. DURATION setting	▼		DUR allows the user to select a RUN Duration time. When a RUN duration period is shown here, it takes precedence over the Timer setting so that any preset start and stop times WILL NOT BE USED. Press ← to enable and then use ▲ and ▼ to select from the following durations: OFF, 5, 10, 15, 30 minutes, 1, 2, 4, 8, 10, 12, 24 hours; confirm with ←.
5. REPEAT setting	▼		Rept allows the user to select up to 49 repeats of the timer settings. Press ← to enable, use ▲ and ▼ to select the number of repeats and then confirm with ←.

6. Logging Pairs of Start/End times	▼		<p>This is the first logging time pair. Press ← to enable / disable each start/end time pair. When a pair is disabled, the L icon will switch off.</p>
7. Begin (start) time	▼		<p>This is the first Begin (start) time: dd-hh:mm. The day setting is the day of the month. When day “ - - ” is indicated, measurement will start on the current day. Press ← to enable, use ▲, ▼, ◀ and ▶ to selecting the Beginning (start) time (dd-hh:mm), and then confirm with ←.</p>
8. End time	▼		<p>This is the first END (stop) time: dd-hh:mm. Press ← to enable, use ▲, ▼, ◀ and ▶ to select the End (stop) time (dd-hh:mm), and then confirm with ←.</p>
9. Review other Start / End Times	▼ repeat	Start / End times	The display steps through 16 sets of Start / End times.

12. Clear the Meter's Memory

Use this option with care – ALL STORED DATA CAN BE DELETED!

Operation	Press	Display Shows	Comments
1. Select the Clear Memory option			Continued from Section 11.6
2. Enable Clear Memory	▼ To toggle steps 1 and 2		Press ← to Clear all stored data while "SurE--" is displayed or press MENU to abort and select another mode of operation.

13. Specifications

Applicable Standards:

- IEC61252, ANSI S1.25 – 1992 for dose meter and sound exposure meters
- IEC 60651 – 1979 Type 2 for sound level meters
- IEC 60804 – 2000 Type 2 for integration sound level meters
- ANSI S1.4 – 1983 Type 2 for sound level meters
- ANSI S1.43 – 1997 for integrating sound level meters

Microphone: 1/2-inch electret condenser microphone with 31-inch integral cable.

Display: Liquid Crystal Display

Measurement Ranges:

Linearity and Indicator ranges at 4 KHz (IEC 60804):

60 – 130: 60 – 130dB (A and C)

70 – 140: 70 – 140dB (A and C)

Peak Ranges:

'C' Frequency Weighted Peak or Linear (Z) Peak over the top 40dB of each range:

60 – 130db: 93 – 133dB Peak

70 – 140db: 103 – 143dB Peak

Frequency Weighting: RMS Detector: 'A' or 'C'; Peak Detector: 'C' or 'Z' (flat or linear).

Response Time Weighting: Fast and Slow (RMS detector).

Exchange Rate (Q): 3, 4, 5 or 6 db

Stabilization Time: 10 seconds after the meter is switched on.

Threshold and Criterion Levels: Threshold Level: 70 – 90dB in 1dB steps; Criterion Level: 80, 84, 85 or 90dB.

Overload Indication (↑): 0.1dB above the selected measurement range.

Under-Range Indication (↓): 1dB below the selected range.

High Level Detector: 115dB

Keypad Lock: ▲ and ► simultaneously to lock; ◀ and ▼ simultaneously to unlock.

Clock: Real-time clock with calendar.

Memory: 20 noise surveys can be stored.

Measurement Control: Measurement Duration may be set to 5, 10, 15, or 30 minutes or 1, 2, 4, 8, 10, 12, or 24 hours.

Timers: Up to 5 timers (automatic start and stop) can be set (maximum up to one month ahead of measurement time). Timer Controlled Start/Stop: Configured using PC software or meter keypad.

Logging: All measurement parameters values can be logged, if the Logging Period is ≥ 1 minute.

Logging Period (LP): Seconds: 1, 2, 5, 10, 15, 20, or 30; Minutes: 1, 2, 5, 10, 15, 20, 30, or 60.

Logging Capacity: The meter can log 14400 total values spread over 20 memory locations (RUNS).

User Setup: Note that one additional user-defined setup configuration can be stored in addition to the factory default configurations shown below.

FACTORY DEFAULT SETTINGS		
Setup	USr	SLM
Measurement Range (dB)	70 – 140	70 – 140
Time Weighting	Fast	Slow
Frequency Weighting	A	A
Peak Frequency Weighting	C	Lin (Z)
Exchange Rate (db)	3	3
Threshold (dB)	80	N/A
Criterion Level (dB)	90	N/A
Allows User to Change Setup	Yes	Yes

Measurement Parameters

DOSE MODE	SLM MODE
DOSE%	SPL
TWA	LEQ
	SEL
	PKZ

PC Interface method: USB interface (meter-to-PC cable included in kit)

Battery: Three (3) 1.5V 'AAA' batteries

Battery life: Typically > 35 hours at room temperature. The instrument will automatically stop measuring and store its data before the battery voltage falls below the cut-off value.

Battery Indicator: Battery voltage level is represented in 4 stages using a displayed battery symbol.

Battery Backup: 3V lithium battery (CR1220)


Operating Temperature & Humidity: 0 – 50°C (32 to 122°F); 10 – 90%RH

Storage Temperature & Humidity: -10 – 60°C (14 to 140°F); 10 – 75%RH

Dimensions: 97(L) × 51(W) × 35(H) mm (3.9 x 2.0 x 1.4")

Weight: Approx. 120g (3.8 oz.) including batteries

CE Certification:

	CE-mark indicates compliance EMC Directive
EMC Emission	<p>EN50081-1 (1992): Generic emission standard. Part 1: Residential, commercial and light industry</p> <p>EN50081-2 (1993): Generic emission standard. Part 2: Industrial environment</p> <p>CISPR22 (1993): Radio disturbance characteristics of information technology equipment. Class B Limits</p> <p>FCC Rules, Part 15: Complies with the Limits for a Class B digital device</p>
EMC Immunity	<p>EN50082-1 (1992): Generic immunity standard. Part 1: Residential, commercial and light industry</p> <p>RF immunity implies that sound level indications of 70dB or greater will be affected by no more than ± 1.5dB</p> <p>EN 50082-2 (1995): Generic immunity standard. Part 2: Industrial environment RF immunity implies that sound level indications of 70dB or greater will be affected by no more than ± 1.5dB</p>

14. Maintenance

Care, Cleaning and Storage

The SL355 is a delicate precision instrument; when handling, storing, or cleaning the instrument, please observe the following:

(a) Storing the Instrument

- Keep the instrument in a dry place.
- For long-term storage, remove the batteries.
- Do not exceed the storage temperature limits: -10 to +60°C (-14 to +140°F)

(b) Cleaning the Instrument

If the instrument casing becomes soiled, wipe it with a cloth that is lightly dampened with water. Do not use abrasive cleansers or solvents. Do not allow moisture to enter the microphone, connectors, or casing.

(c) Handling the instrument

- Do not attempt to remove the microphone grid; the microphone can become easily damaged if it is opened.
- Do not attempt to open the instrument; there are no user-serviceable parts inside. If the instrument requires service, please contact the point of sale.
- Do not allow the instrument to come in contact with moisture.
- Never mix battery types.
- Never mix charged and discharged batteries.
- Do not allow fully discharged batteries to remain inside the instrument.
- Protect the instrument from impact.

15. Software Installation and Operation

The supplied CD-ROM includes the SL355 Software program, driver, and the Software Help Guide. Please refer to the Software Help Guide for detailed instructions on the use of the supplied software. Visit the Extech web site (www.extech.com) to check the latest version of the software; download a newer version if necessary.

16. Glossary of Terms

% DOSE

The unit of measure, % DOSE, is used to quantify noise exposure measured during a work shift. 100% dose is the maximum allowable noise exposure in accordance with OSHA, MSHA, DOD, ACGIH, and ISO standards. Most standards specify Criterion Level, Exchange Rate, Response Time, and Frequency weighting for the dosimeter.

CRITERION LEVEL

To take an on-site noise exposure survey in accordance with standards such as OSHA and MSHA, the dosimeter's Criterion Level must first be set. The 100% DOSE parameter discussed above is determined by the following equation: 100% DOSE = Criterion Level for 8 hours. Each country has a unique Criterion Level (most countries, including the U.S., use 90dB). The Criterion level is selectable (80, 84, 85, or 90dB).

EXCHANGE RATE

Exchange rate is best illustrated by example, as follows: Since 100% DOSE = Criterion Level for 8 hours, a person would receive 50% DOSE in 4 hours if the noise level equals the Criterion Level setting. Now consider a Criterion Level of 90dB, a noise measurement of 95dB (5dBA higher than the Criterion Level), and an Exchange Rate of 5dB; in this example a 100% DOSE would be received in only 4 hours. This is because with a 5dB Exchange Rate, a 5dB increase in sound level is considered a doubling of the DOSE. Other Exchange Rates can be selected (3, 4, 5, or 6 db). Refer to the local regulations or standards for exchange rate, criterion level and threshold settings.

FAST (F) SLOW (S) RESPONSE TIME

Set the response time to Fast (F) to capture quick bursts of sound such as discharging firearms, fireworks, hammering, and other impulse noises. Use the Slow (S) setting if the noise under test is more of a continuous drone or background din. The Slow setting is typically specified by OSHA and MSHA standards for use in noise surveys.

THRESHOLD LEVEL

The threshold level is the sound level at which the meter begins to integrate noise into the exposure test. For example, if the threshold level is set to 85dB, the meter will integrate all noise that equals or exceeds 85 dB. Sound levels below this threshold would not be included in the dose calculation.


TIME WEIGHTED AVERAGE (TWA)

TWA is the 8-hour projected average sound level based on data collected from the start of the test up to the time the TWA is checked. TWA sound exposure is displayed in Pa²h on the SL355

PEAK

When sound levels above 140dB are present, the meter displays the PK (and the 'C' or 'Z' weighting) symbols. The 'C' icon indicates 'C' frequency weighting and the 'Z' icon indicates flat (or linear) peak where no frequency weighting is applied.

HIGH LEVEL INDICATOR

When sound levels above 115dB are present the meter displays the headphone  symbol.